

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Superfund Records Center

STE: Shpack

BLANK: 13.1

WATER: 209660

Client Sample ID: 59680 **TRIP VOA BLANK**
Date of Collection: 8/25/03
Date of Extraction: 8/27/03
Date of Analysis: 8/27/03
Dry Weight Extracted: N/A
Wet Weight Extracted: N/A

Lab Sample ID: AA33627
Matrix: DW
Volume Purged: 5 mL
Percent Solids: N/A
Extract Dilution: 1
pH: 5.2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

SDMS DocID 000209660



124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	100	95 - 112
Toluene-D8	102	94 - 109
1,4-Bromofluorobenzene	102	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59681 **68 Union Street**

Date of Collection: 8/26/03

Date of Extraction: 8/27/03

Date of Analysis: 8/27/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33628

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: --<2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

74-87-3	Chloromethane	ND	0.50	
124-48-1	Dibromochloromethane	ND	0.50	
74-95-3	Dibromomethane	ND	0.50	
75-71-8	Dichlorodifluoromethane	ND	0.50	
60-29-7	Ethyl Ether	ND	0.50	
100-41-4	Ethylbenzene	ND	0.50	
87-68-3	Hexachlorobutadiene	ND	0.50	
98-82-8	Isopropylbenzene	ND	0.50	
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0	
1634-04-4	Methyl-t-butyl-ether	ND	0.50	
75-09-2	Methylene Chloride	ND	0.50	
104-51-8	N-Butylbenzene	ND	0.50	
103-65-1	N-Propylbenzene	ND	0.50	
91-20-3	Naphthalene	ND	0.50	--
95-47-6	Ortho Xylene	ND	0.50	
99-87-6	Para-Isopropyltoluene	ND	0.50	
135-98-8	Sec-Butylbenzene	ND	0.50	
100-42-5	Styrene	ND	0.50	
98-06-6	Tert-Butylbenzene	ND	0.50	
127-18-4	Tetrachloroethylene	ND	0.50	
109-99-9	Tetrahydrofuran	ND	5.0	
108-88-3	Toluene	ND	0.50	
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50	
79-01-6	Trichloroethylene	ND	0.50	
75-69-4	Trichlorofluoromethane	ND	0.50	
108-05-4	Vinyl Acetate	ND	0.50	
75-01-4	Vinyl Chloride	ND	0.50	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	
10061-01-5	cis-1,3-dichloropropene	ND	0.50	
10061-02-6	t-1,3-Dichloropropene	ND	0.50	

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	100	94 - 109
1,4-Bromofluorobenzene	100	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59682 **106 Peckham Street**

Date of Collection: 8/26/03

Date of Extraction: 8/27/03

Date of Analysis: 8/27/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33629

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: -- <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50	
74-95-3	Dibromomethane	ND	0.50	
75-71-8	Dichlorodifluoromethane	ND	0.50	
60-29-7	Ethyl Ether	ND	0.50	
100-41-4	Ethylbenzene	ND	0.50	
87-68-3	Hexachlorobutadiene	ND	0.50	
98-82-8	Isopropylbenzene	ND	0.50	
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0	
1634-04-4	Methyl-t-butyl-ether	ND	0.50	
75-09-2	Methylene Chloride	ND	0.50	
104-51-8	N-Butylbenzene	ND	0.50	
103-65-1	N-Propylbenzene	ND	0.50	
91-20-3	Naphthalene	ND	0.50	-
95-47-6	Ortho Xylene	ND	0.50	
99-87-6	Para-Isopropyltoluene	ND	0.50	
135-98-8	Sec-Butylbenzene	ND	0.50	
100-42-5	Styrene	ND	0.50	
98-06-6	Tert-Butylbenzene	ND	0.50	
127-18-4	Tetrachloroethylene	ND	0.50	
109-99-9	Tetrahydrofuran	ND	5.0	
108-88-3	Toluene	ND	0.50	
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50	
79-01-6	Trichloroethylene	ND	0.50	
75-69-4	Trichlorofluoromethane	ND	0.50	
108-05-4	Vinyl Acetate	ND	0.50	
75-01-4	Vinyl Chloride	ND	0.50	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	
10061-01-5	cis-1,3-dichloropropene	ND	0.50	
10061-02-6	t-1,3-Dichloropropene	ND	0.50	

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	100	95 - 112
Toluene-D8	102	94 - 109
1,4-Bromofluorobenzene	100	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59683 **100 Peckham Street**

Date of Collection: 8/26/03

Date of Extraction: 8/27/03

Date of Analysis: 8/27/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33630

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

74-87-3	Chloromethane	ND	0.50
124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	98	95 - 112
Toluene-D8	104	94 - 109
1,4-Bromofluorobenzene	102	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59684 **77 Peckham Street**

Date of Collection: 8/26/03

Date of Extraction: 8/27/03

Date of Analysis: 8/27/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33631

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

74-85-3	Chloromethane	ND	0.50	
124-48-1	Dibromochloromethane	ND	0.50	
74-95-3	Dibromomethane	ND	0.50	
75-71-8	Dichlorodifluoromethane	ND	0.50	
60-29-7	Ethyl Ether	ND	0.50	
100-41-4	Ethylbenzene	ND	0.50	
87-68-3	Hexachlorobutadiene	ND	0.50	
98-82-8	Isopropylbenzene	ND	0.50	
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0	
1634-04-4	Methyl-t-butyl-ether	22	0.50	
75-09-2	Methylene Chloride	ND	0.50	
104-51-8	N-Butylbenzene	ND	0.50	
103-65-1	N-Propylbenzene	ND	0.50	
91-20-3	Naphthalene	ND	0.50	--
95-47-6	Ortho Xylene	ND	0.50	
99-87-6	Para-Isopropyltoluene	ND	0.50	
135-98-8	Sec-Butylbenzene	ND	0.50	
100-42-5	Styrene	ND	0.50	
98-06-6	Tert-Butylbenzene	ND	0.50	
127-18-4	Tetrachloroethylene	ND	0.50	
109-99-9	Tetrahydrofuran	ND	5.0	
108-88-3	Toluene	ND	0.50	
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50	
79-01-6	Trichloroethylene	ND	0.50	
75-69-4	Trichlorofluoromethane	ND	0.50	
108-05-4	Vinyl Acetate	ND	0.50	
75-01-4	Vinyl Chloride	ND	0.50	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	
10061-01-5	cis-1,3-dichloropropene	ND	0.50	
10061-02-6	t-1,3-Dichloropropene	ND	0.50	

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	102	94 - 109
1,4-Bromofluorobenzene	104	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59685 **77 Peckham Street**
Date of Collection: 8/26/03 **(field duplicate)**
Date of Extraction: 8/27/03
Date of Analysis: 8/27/03
Dry Weight Extracted: N/A
Wet Weight Extracted: N/A

Lab Sample ID: AA33632
Matrix DW
Volume Purged: 5 mL
Percent Solids: N/A
Extract Dilution: 1
pH: ≤2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

74-87-3	Chloromethane	ND	0.50
124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	23	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	104	94 - 109
1,4-Bromofluorobenzene	102	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59686 **70 Peckham Street**

Date of Collection: 8/26/03

Date of Extraction: 8/27/03

Date of Analysis: 8/27/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33633

Matrix: DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50	
74-95-3	Dibromomethane	ND	0.50	
75-71-8	Dichlorodifluoromethane	ND	0.50	
60-29-7	Ethyl Ether	ND	0.50	
100-41-4	Ethylbenzene	ND	0.50	
87-68-3	Hexachlorobutadiene	ND	0.50	
98-82-8	Isopropylbenzene	ND	0.50	
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0	
1634-04-4	Methyl-t-butyl-ether	ND	0.50	
75-09-2	Methylene Chloride	ND	0.50	
104-51-8	N-Butylbenzene	ND	0.50	
103-65-1	N-Propylbenzene	ND	0.50	
91-20-3	Naphthalene	ND	0.50	--
95-47-6	Ortho Xylene	ND	0.50	
99-87-6	Para-Isopropyltoluene	ND	0.50	
135-98-8	Sec-Butylbenzene	ND	0.50	
100-42-5	Styrene	ND	0.50	
98-06-6	Tert-Butylbenzene	ND	0.50	
127-18-4	Tetrachloroethylene	ND	0.50	
109-99-9	Tetrahydrofuran	ND	5.0	
108-88-3	Toluene	ND	0.50	
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50	
79-01-6	Trichloroethylene	ND	0.50	
75-69-4	Trichlorofluoromethane	ND	0.50	
108-05-4	Vinyl Acetate	ND	0.50	
75-01-4	Vinyl Chloride	ND	0.50	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	
10061-01-5	cis-1,3-dichloropropene	ND	0.50	
10061-02-6	t-1,3-Dichloropropene	ND	0.50	

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	104	94 - 109
1,4-Bromofluorobenzene	104	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59687 **14 North Worcester Street**
Date of Collection: 8/26/03
Date of Extraction: 8/28/03
Date of Analysis: 8/28/03
Dry Weight Extracted: N/A
Wet Weight Extracted: N/A

Lab Sample ID: AA33634
Matrix: DW
Volume Purged: 5 mL
Percent Solids: N/A
Extract Dilution: 1
pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

74-81-2	Chloromethane	ND	0.50
124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	106	94 - 109
1,4-Bromofluorobenzene	104	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59688 **35 Maple Street**
Date of Collection: 8/26/03
Date of Extraction: 8/28/03
Date of Analysis: 8/28/03
Dry Weight Extracted: N/A
Wet Weight Extracted: N/A

Lab Sample ID: AA33635
Matrix: DW
Volume Purged: 5 mL
Percent Solids: N/A
Extract Dilution: 1
pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	104	95 - 112
Toluene-D8	104	94 - 109
1,4-Bromofluorobenzene	102	73 - 113

Comments:

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59689 **36 Maple Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33636

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	100	95 - 112
Toluene-D8	106	94 - 109
1,4-Bromofluorobenzene	104	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59690 **82 Maple Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33637

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	100	95 - 112
Toluene-D8	104	94 - 109
1,4-Bromofluorobenzene	102	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59691 **83 Maple Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33638

Matrix: DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	106	94 - 109
1,4-Bromofluorobenzene	106	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59692 **95 Maple Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33639

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MJBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	106	94 - 109
1,4-Bromofluorobenzene	104	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59693 **94 Maple Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33640

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: ≤2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	ND	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	ND	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	100	95 - 112
Toluene-D8	102	94 - 109
1,4-Bromofluorobenzene	100	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.

US ENVIRONMENTAL PROTECTION AGENCY
NEW ENGLAND LABORATORY

Shpack Landfill - Norton MA

VOAs in Drinking Water

Client Sample ID: 59694 **59 Union Street**

Date of Collection: 8/26/03

Date of Extraction: 8/28/03

Date of Analysis: 8/28/03

Dry Weight Extracted: N/A

Wet Weight Extracted: N/A

Lab Sample ID: AA33641

Matrix DW

Volume Purged: 5 mL

Percent Solids: N/A

Extract Dilution: 1

pH: <2

CAS Number	Compound	Concentration ug/L	RL ug/L	Qualifier
630-20-6	1,1,1,2-Tetrachloroethane	ND	0.50	
71-55-6	1,1,1-Trichloroethane	ND	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	
76-13-1	1,1,2-Trichloro-1,2,2-Trifluoroeth	ND	0.50	
79-00-5	1,1,2-Trichloroethane	ND	0.50	
75-35-4	1,1-Dichloroethylene	ND	0.50	
563-58-6	1,1-Dichloropropene	ND	0.50	
75-34-3	1,1-dichloroethane	ND	0.50	
87-61-6	1,2,3-Trichlorobenzene	ND	0.50	
96-18-4	1,2,3-Trichloropropane	ND	0.50	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	
96-12-8	1,2-Dibromo-3-Chloropropane	ND	0.50	
106-93-4	1,2-Dibromoethane	ND	0.50	
95-50-1	1,2-Dichlorobenzene	ND	0.50	
107-06-2	1,2-Dichloroethane	ND	0.50	
78-87-5	1,2-Dichloropropane	ND	0.50	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	
142-28-9	1,3-Dichloropropane	ND	0.50	
541-73-1	1,3-dichlorobenzene	ND	0.50	
106-46-7	1,4-Dichlorobenzene	ND	0.50	
590-20-7	2,2-Dichloropropane	ND	0.50	
78-93-3	2-Butanone (MEK)	ND	0.50	
95-49-8	2-Chlorotoluene	ND	0.50	
591-10-6	2-Hexanone	ND	0.50	
106-43-4	4-Chlorotoluene	ND	0.50	
108-10-1	4-Methyl-2-Pentanone(MIBK)	ND	0.50	
67-64-1	Acetone	ND	0.50	
107-13-1	Acrylonitrile	ND	5.0	
71-43-2	Benzene	ND	0.50	
108-86-1	Bromobenzene	ND	0.50	
74-97-5	Bromochloromethane	ND	0.50	
75-27-4	Bromodichloromethane	ND	0.50	
75-25-2	Bromoform	ND	0.50	
74-83-9	Bromomethane	ND	0.50	
75-15-0	Carbon Disulfide	ND	0.50	
56-23-5	Carbon tetrachloride	ND	0.50	
108-90-7	Chlorobenzene	ND	0.50	
75-00-3	Chloroethane	ND	0.50	
67-66-3	Chloroform	2.1	0.50	

124-48-1	Dibromochloromethane	ND	0.50
74-95-3	Dibromomethane	ND	0.50
75-71-8	Dichlorodifluoromethane	ND	0.50
60-29-7	Ethyl Ether	ND	0.50
100-41-4	Ethylbenzene	ND	0.50
87-68-3	Hexachlorobutadiene	ND	0.50
98-82-8	Isopropylbenzene	ND	0.50
108-38-3 / 106-4	Meta/Para Xylene	ND	1.0
1634-04-4	Methyl-t-butyl-ether	1.8	0.50
75-09-2	Methylene Chloride	ND	0.50
104-51-8	N-Butylbenzene	ND	0.50
103-65-1	N-Propylbenzene	ND	0.50
91-20-3	Naphthalene	ND	0.50
95-47-6	Ortho Xylene	ND	0.50
99-87-6	Para-Isopropyltoluene	ND	0.50
135-98-8	Sec-Butylbenzene	ND	0.50
100-42-5	Styrene	ND	0.50
98-06-6	Tert-Butylbenzene	ND	0.50
127-18-4	Tetrachloroethylene	ND	0.50
109-99-9	Tetrahydrofuran	ND	5.0
108-88-3	Toluene	ND	0.50
156-60-5	Trans-1,2-Dichloroethylene	ND	0.50
79-01-6	Trichloroethylene	ND	0.50
75-69-4	Trichlorofluoromethane	ND	0.50
108-05-4	Vinyl Acetate	ND	0.50
75-01-4	Vinyl Chloride	ND	0.50
156-59-2	cis-1,2-Dichloroethylene	ND	0.50
10061-01-5	cis-1,3-dichloropropene	ND	0.50
10061-02-6	t-1,3-Dichloropropene	ND	0.50

Surrogate Compounds	Recoveries (%)	QC Ranges
1,2-Dichlorobenzene-d4	102	95 - 112
Toluene-D8	102	94 - 109
1,4-Bromofluorobenzene	100	73 - 113

Comments: The RL for acrylonitrile and tetrahydrofuran have been raised to 5.0 ppb to reflect points omitted from the initial calibration.